



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/590,229

06/28/2007

Elizabeth Rajan

14283.0013USWO

6747

23552 7590 06/22/2009
MERCHANT & GOULD PC
P.O. BOX 2903
MINNEAPOLIS, MN 55402-0903

EXAMINER

TREYGER, ILYA Y

ART UNIT

PAPER NUMBER

3761

MAIL DATE

DELIVERY MODE

06/22/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/590,229	Applicant(s) RAJAN ET AL.	
	Examiner ILYA Y. TREYGER	Art Unit 3761	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 31 March 2009.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-139 is/are pending in the application.
- 4a) Of the above claim(s) 9 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-8 and 10-13 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 21 August 2006 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

1. Claims 1, 5, and 7 are amended.
2. Claim 9 is canceled.
3. Claims 10-13 are new.
4. Claims 1-8 and 10-13 are examined on the merits.

Response to Arguments

5. Applicant's arguments filed 03/31/2009 have been fully considered but they are not persuasive.
6. With regard to claim 1, Applicants argue that Rockey does not disclose the claimed invention because the reference does not disclose the device with a bower extension and comprising permeable material.

However, Stack teaches a gastric bypass comprising material that allows the substances to pass from the reservoirs into the surrounding tissue (page 2, [0039], lines 9-11) and a bowel extension B (Fig. 4B).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to supply the apparatus of Rockey with the permeable material, an esophageal extension and a bowel extension, as taught by Stack in order to utilize the bypass of the conventionally known type.

7. With respect to claims 5 and 6, Applicants argue that the claimed invention has not been disclosed by the reference because it does not disclose a plurality of adjacent subchambers.

However, it would have been obvious to one having ordinary skill in the art at the time the invention was made to make the expandable chamber in form of plurality of

Art Unit: 3761

adjacent subchambers such as a quasi-geodesic pattern in order to increase the square area of the chamber, since it has been held that mere multiplication of the essential working parts of a device involves only routine skill in the art.

8. With respect to claims 3 and 4, Applicant's arguments are substantially identical to arguments discussed above.

Claim Rejections - 35 USC § 112

9. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

10. Claim 7 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. It is indefinite term of the "asymmetric toroidal chamber", since in accordance with the definition, toroid is a surface generated by a closed curve rotating about, but not intersecting or containing, an axis in its own plane, that reads on is symmetric by the definition.

11.

Claim Rejections - 35 USC § 103

12. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Art Unit: 3761

13. The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

14. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

15. Claims 1-8 and 10-13 are rejected under 35 U.S.C. 103(a) as being unpatentable over Rockey (US 4,763,653) in view of Stack et al. (US 2003/0199991).

16. In Re claim 1, Rockey discloses a sleeve unit insertable in a natural body vessel (Abstract, lines 2, 3) capable of being used as a gastro-intestinal device (Col. 2, lines6-8), comprising:

a sleeve 51 (Fig. 9) that is a gastric bypass having an inlet A (Fig. 9) and outlet B (Fig. 9); and

an expandable inflatable balloon 50 (Fig. 9) attached to the gastric bypass (Col. 5, lines 64-68).




FIG. 9

Rockey does not expressly disclose the gastric bypass comprising a permeable material; an esophageal extension; and a bowel extension.

Stack teaches a gastric bypass comprising material that allows the substances to pass from the reservoirs into the surrounding tissue (page 2, [0039], lines 9-11); an esophageal extension A (Fig. 17C); and a bowel extension B (Fig. 4B).




FIG. 4B




FIG. 17C

It would have been obvious to one having ordinary skill in the art at the time the invention was made to supply the apparatus of Rockey with the permeable material, an esophageal extension and a bowel extension, as taught by Stack in order to utilize the bypass of the conventionally known type.

Where a claimed improvement on a device or apparatus is no more than "the simple substitution of one known element for another or the mere application of a known technique to a piece of prior art ready for improvement," the claim is unpatentable under 35 U.S.C. § 103(a). *Ex Parte Smith*, 83 USPQ.2d 1509, 1518-19 (BPAI, 2007)(citing *KSR v. Teleflex*, 127 S.Ct. 1727, 1740, 82 USPQ2d 1385, 1396 (2007)). Accordingly, Applicant claims a combination that only unites old elements with no change in the respective functions of those old elements, and the combination of those elements yields predictable results; absent persuasive evidence that the modifications necessary to effect the combination of elements is uniquely challenging or difficult for one of ordinary skill in the art, the claim is unpatentable as obvious under 35 U.S.C. § 103(a).

Art Unit: 3761

Ex Parte Smith, 83 USPQ.2d at 1518-19 (BPAI, 2007)(citing *KSR*, 127 S.Ct. at 1740, 82 USPQ2d at 1396). Accordingly, since the applicant[s] have submitted no persuasive evidence that the combination of the above elements is uniquely challenging or difficult for one of ordinary skill in the art, the claim is unpatentable as obvious under 35 U.S.C. § 103(a) because it is no more than the predictable use of prior art elements according to their established functions resulting in the simple substitution of one known element for another or the mere application of a known technique to a piece of prior art ready for improvement.

17. In Re claims 2 and 8, Rockey discloses the gastric bypass comprising the flaccid gastric bag, since the sleeve (sheath) is disclosed as formed of flexible material (Col. 3, line 33).

18. In Re claims 3 and 10, Rockey discloses the invention discussed above, but does not expressly disclose the device having ribs facilitating collapse of the tube in the absence of compressive forces.

Stack teaches the gastric bypass comprising radially extending tabs 13 (Fig. 16B) that are ribs (claim 10) facilitating collapse of the tube in the absence of compressive forces (page 3, [0045], lines 1-6).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the device of Rockey with the tabs, as taught by Stack in order to facilitate collapse of the tube (page 3, [0045], line 2).

19. In Re claims 4 and 13, Rockey discloses the invention discussed above, but does not expressly disclose the device formed of a polymeric sleeve with nitinol struts embedded in the sleeve material.

Art Unit: 3761

Stack teaches the device formed of a polymeric sleeve with nitinol struts embedded in the sleeve material (See page 4, [0060], lines 8-10), wherein the wire is disposed helically (claim 13).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the device of Rockey with the nitinol struts, as taught by Stack in order to provide the device with the additional rigidity.

20. Rockey discloses an implantable satiation device (page 1, [0005], lines 1-4) that is a gastro-intestinal device, comprising:

a tube 10c (Fig. 13) that is a gastric bypass having an inlet A (Fig. 13) and outlet B (Fig. 13); and

an expandable inflatable reservoir 54 (Fig. 13) attached to the gastric bypass (See page 4, [0059] and [0060]).

Rockey does not expressly disclose the device wherein the expandable chamber is made in form of plurality of adjacent subchambers.

It would have been obvious to one having ordinary skill in the art at the time the invention was made to make the expandable chamber in form of plurality of adjacent subchambers such as a quasi-geodesic pattern (claim 6) in order to increase the square area of the chamber, since it has been held that mere multiplication of the essential working parts of a device involves only routine skill in the art.

21. In Re claim 7, Stack discloses an implantable satiation device (page 1, [0005], lines 1-4) that is a gastro-intestinal device, comprising:

Art Unit: 3761

a tube 10c (Fig. 13) that is a gastric bypass having an inlet A (Fig. 13) and outlet B (Fig. 13); and

Rockey discloses a sleeve unit insertable in a natural body vessel (Abstract, lines 2, 3) capable of being used as a gastro-intestinal device (Col. 2, lines 6-8), comprising:

a sleeve 51 (Fig. 9) that is a gastric bypass having an inlet A (Fig. 13) and outlet B (Fig. 9); and

an expandable toroidal inflatable balloon 50 (Fig. 9) attached to the gastric bypass (Col. 5, lines 64-68) fully capable of taking the asymmetric form, since it is made of flexible material .

22. In Re claims 11 and 12, Rockey in view of Stack disclose the invention discussed above, as applied to claim 3, but do not expressly disclose the particular disposition character of the ribs.

Since Applicant did not state that such disposition character serves any specific purpose or performs any specific function other than the function disclosed in Stack, i.e. facilitating collapse of the tube, such disposition character is the matter of the obvious design choice, and therefore it would have been obvious to those skilled in the art at the time the invention was made to dispose ribs helically or longitudinally as an obvious design choice, and as such it does not impact the patentability of claims.

Conclusion

23. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

Art Unit: 3761

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to ILYA Y. TREYGER whose telephone number is (571)270-3217. The examiner can normally be reached on 7:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Tatyana Zalukaeva can be reached on 571-272-1115. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Application/Control Number: 10/590,229

Page 11

Art Unit: 3761

/Ilya Y Treyger/

Examiner, Art Unit 3761

/Michele Kidwell/

Primary Examiner, Art Unit 3761